



**Technology Transfer Seminar**  
**Practical Solutions for Applying**  
**BS EN 14181: 2004**  
**Bretby, 21<sup>st</sup> September 2006**

**The Importance of QAL3**  
**Paul Mudway**  
**Source Testing Association**

The text is overlaid on a background image of industrial testing equipment, including a control panel with a screen and various cables.

## The importance of QAL3

- QAL3 - General
- Statistical Process Control (SPC)
- Calculation of  $S_{AMS}$
- Example control charts
- Reporting requirements
- Summary
- Sources of information

The Importance of QAL3. Presentation. TTS 21/9/06  
Version 1.0, Slide 3



## QAL3 – general

- QAL3 is a procedure to maintain and demonstrate the quality of the AMS during ongoing operation.
- The purpose of QAL3 is to detect drift and changes in precision in the AMS by performing regular checks of the zero and span readings.
- These checks may be made using reference materials or by using surrogate methods.

The Importance of QAL3. Presentation. TTS 21/9/06  
Version 1.0, Slide 4



## QAL3 – general (cont.)

- In order to determine whether any changes in the zero and span values are due to actual drift, or whether they are caused by random deviations, zero and span values are plotted on control charts and compared to allowable variations expressed as standard deviations, or  $S_{AMS}$  in EN 14181.

The Importance of QAL3. Presentation. TTS 21/9/06  
Version 1.0, Slide 5



## QAL3 – general (cont.)

Do not undertake QAL2 tests until QAL3 procedures are in operation (strongly recommended) or are ready to operate immediately after QAL2 tests are completed.

The Importance of QAL3. Presentation. TTS 21/9/06  
Version 1.0, Slide 6



## Who carries out QAL3?

- The implementation and performance of QAL3 procedures are the responsibility of the plant operator.
- It is also the responsibility of the plant operator to ensure that the AMS is operating within the valid calibration range.

The Importance of QAL3. Presentation. TTS 21/9/06  
Version 1.0, Slide 7



## Statistical process control (SPC)

- Statistical process control (SPC) is a method for achieving quality control in manufacturing processes.
- Control charts or process behaviour charts are widely used to assess the nature of variation in a process and to facilitate forecasting and management.
- EN 14181 suggests the use of two types of control chart to detect drift and changes in precision in the AMS: Shewhart charts and CUSUM control charts. Other types may be used.

The Importance of QAL3. Presentation. TTS 21/9/06  
Version 1.0, Slide 8



## Statistical process control (SPC) (cont.)

- Shewhart charts are good at detecting large changes in the process mean or variance.
- CUSUM charts, while not as intuitive and simple to operate as Shewhart charts, are more efficient in detecting small drifts in the mean of a process.

The Importance of QAL3. Presentation. TTS 21/9/06  
Version 1.0, Slide 9



## Why do we need control charts?

- Frequent adjustments to AMS are often not needed.
- Frequent adjustments can degrade accuracy.
- Key objective is to determine drift.

The Importance of QAL3. Presentation. TTS 21/9/06  
Version 1.0, Slide 10



## What is $S_{AMS}$ ?

- $S_{AMS}$  is the estimated uncertainty of the AMS, expressed as a standard deviation.
- Both Shewhart and CUSUM control charts use multiples of  $S_{AMS}$  to determine if changes in zero and span values are due to actual drift, or to random deviations due to variations in voltage, ambient temperature etc.
- Shewhart and CUSUM charts use different multiples of  $S_{AMS}$  as warning levels and alarm levels.

The Importance of QAL3. Presentation. TTS 21/9/06  
Version 1.0, Slide 11



## Calculation of $S_{AMS}$

The standard deviation  $S_{AMS}$  is calculated by:

$$S_{AMS} = \sqrt{(u_{temp})^2 + (u_{volt})^2 + (u_{pres})^2 + (u_{others})^2}$$

Where

- $u_{temp}$  is the uncertainty from variations in ambient temperature
- $u_{volt}$  is the uncertainty from variations in voltage
- $u_{pres}$  is the uncertainty from variations in stack gas pressure for in situ AMS
- $u_{others}$  is the uncertainty from other sources, e.g. cross-sensitivity to other determinands, detection limit etc.

The Importance of QAL3. Presentation. TTS 21/9/06  
Version 1.0, Slide 12



## Target values for control charts

- Shewhart and CUSUM approaches require target values to be chosen for span and zero.
- The target values (C) could be the concentration of the span or zero gas used for calibration checks, or a surrogate calibration method.

The Importance of QAL3. Presentation. TTS 21/9/06  
Version 1.0, Slide 13



## What span gas concentration to use?

- It is not necessary to use calibration gases supplied by a manufacturer accredited to the EN ISO/IEC 17025 standard, but the gases must be stable.
- As a “rule of thumb”, the span gas concentration should be about 75% of 1.5 times the half-hourly ELV set for the process.
- If half-hourly ELV is not set, choose a concentration approximately equal to the normal, average stack gas concentration.

The Importance of QAL3. Presentation. TTS 21/9/06  
Version 1.0, Slide 14



## Shewhart control charts

- There are many types of Shewhart control charts including average, range, standard deviation, individuals, moving range etc.
- The simplest chart which may be plotted is one for individual measurements, known as an individuals chart.

The Importance of QAL3. Presentation. TTS 21/9/06  
Version 1.0, Slide 15




## Shewhart control charts (cont.)

- The limits are set using multiples of the uncertainty of the AMS or  $S_{AMS}$  for the applicable certified range.
- For Shewhart individuals charts, with a sample number of one, i.e. one calibration, it is usual to use warning and alarm limits of  $2 \times S_{AMS}$  (approximately 95% CI) and  $3 \times S_{AMS}$  (approximately 99% CI) respectively.

The Importance of QAL3. Presentation. TTS 21/9/06  
Version 1.0, Slide 16



## Example template for Shewhart individuals charts – span chart




The Importance of QAL3. Presentation. TTS 21/9/06  
Version 1.0, Slide 17



## Example Shewhart individuals chart

Shewhart chart for span




The Importance of QAL3. Presentation. TTS 21/9/06  
Version 1.0, Slide 18



## Example Shewhart individuals chart

Shewhart chart for zero



The Importance of QAL3. Presentation. TTS 21/9/06  
Version 1.0, Slide 19



## Shewhart control charts (cont.)

- EN 14181 requires the operator to intervene when:
  - One or more data points are beyond one of the alarm limits
  - Three consecutive data points are beyond one of the warning limits
  - Four points among five consecutive ones are beyond half the alarm limits for span
  - Eight consecutive points are on the same side of the mean
  - Six consecutive points are either increasing or decreasing

The Importance of QAL3. Presentation. TTS 21/9/06  
Version 1.0, Slide 20



## CUSUM charts

- The CUSUM or cumulative sum chart uses all of the data and is therefore a more sensitive way to detect slight changes in the mean.
- If a target  $C$  is being considered, then the operating principle of the chart is to calculate the difference between each new value and value  $C$ , and to add this to a cumulative sum.

The Importance of QAL3. Presentation. TTS 21/9/06  
Version 1.0, Slide 21



## CUSUM charts (cont.)

- A CUSUM chart is essentially a running total of deviations from some pre-selected reference value.
- It is sensitive in detecting changes in the mean.
- Any change in the mean, and the extent of the change, is indicated visually by a change in the slope of the graph:
  - the CUSUM slope is upwards, observations are above target
  - the CUSUM slope is downwards, observations are below target
  - the CUSUM slope is horizontal, observations are on target
  - the CUSUM slope changes, the observations are changing level
  - the absolute value of the CUSUM score has little meaning

The Importance of QAL3. Presentation. TTS 21/9/06  
Version 1.0, Slide 22



## CUSUM charts (cont.)


- If  $\mu_0$  is the target value for the process mean,  $\bar{x}_j$  is the average of the  $j^{\text{th}}$  sample, then the CUSUM chart is formed by plotting the quantity:

$$C_i = \sum_{j=1}^i (\bar{x}_j - \mu_0)$$

The Importance of QAL3. Presentation. TTS 21/9/06  
Version 1.0, Slide 23




## Example CUSUM chart for span drift



The Importance of QAL3. Presentation. TTS 21/9/06  
Version 1.0, Slide 24



## Example CUSUM chart for zero drift



The Importance of QAL3. Presentation. TTS 21/9/06  
Version 1.0, Slide 25



## Design and use of control charts according to Annex C of EN 14181

- Annex C of EN 14181 contains specifications for the design and use of control charts for AMS drift and precision.
- In this approach the amount of AMS drift and decrease in precision is compared to the uncertainty of the measured value.

The Importance of QAL3. Presentation. TTS 21/9/06  
Version 1.0, Slide 26



## Design and use of control charts according to Annex C of EN 14181 (cont.)

- If either  $\Sigma(\text{pos})_t$  or  $\Sigma(\text{neg})_t$  exceed the alarm limits then adjustment of the AMS is required.
- If  $s_t$  exceeds the alarm limits the instrument manufacturer shall be contacted.

The Importance of QAL3. Presentation. TTS 21/9/06  
Version 1.0, Slide 27



## Design and use of control charts according to Annex C of EN 14181 (cont.)


- Annex C of BS EN 14181 specifies  $0.501$  and  $2.85 \times S_{\text{AMS}}$  for drift and  $1.85$  and  $6.9 \times S_{\text{AMS}}^2$  for precision.
- In practice limits will differ depending on the performance of the analyser.

The Importance of QAL3. Presentation. TTS 21/9/06  
Version 1.0, Slide 28



## Example EN 14181 control charts – span drift

QAL3 control chart for span drift according to Annex C of EN 14181




The Importance of QAL3. Presentation. TTS 21/9/06  
Version 1.0, Slide 29



## Example EN 14181 control charts – zero drift

QAL3 control chart for zero drift according to Annex C of EN 14181



The Importance of QAL3. Presentation. TTS 21/9/06  
Version 1.0, Slide 30



## QAL3 reporting requirements

- AMS details – monitoring approach and technique, operating range, make & model.
- AMS changes – details of changes in make, model & serial numbers.
- Service visit records – routine maintenance.

The Importance of QAL3. Presentation. TTS 21/9/06  
Version 1.0, Slide 31



## QAL3 reporting requirements (cont.)

- Manufacturer's call out records.
- Operator's routine maintenance and corrective actions.
- QAL3 baseline re-sets – summary.
- Zero & span drift plots.
- Zero & span drift tabulation.

The Importance of QAL3. Presentation. TTS 21/9/06  
Version 1.0, Slide 32



## Summary

- Do not carry out QAL2 tests until QAL3 procedures are in operation.
- Shewhart and CUSUM control charts are recommended – other types may be used.
- Shewhart charts are good at detecting large changes in the process mean or variance.
- CUSUM charts are not as simple or intuitive as Shewhart charts, but are better at detecting small drifts.
- The procedure specified in Annex C of EN 14181 is not a simple CUSUM. It can be used to determine if, and by how much, the AMS needs to be externally adjusted.

The Importance of QAL3. Presentation. TTS 21/9/06  
Version 1.0, Slide 33



## Sources of information

- BS EN 14181: 2004.
- Technical Guidance Note M20.
- Method Implementation Document MID 14181.
- Method Implementation Document MID 13284-2.
- BS 7785: 1994 – Shewhart Control Charts.
- BS 5703-1: 2003 – CUSUM Control Charts.

The Importance of QAL3. Presentation. TTS 21/9/06  
Version 1.0, Slide 34





**THE SOURCE TESTING ASSOCIATION**

**[WWW.S-T-A.ORG](http://WWW.S-T-A.ORG)**