



monitoring & control for your environment



# How To Achieve Compliance WID/EN14181

## ANNEX 2 RELEASES INTO AIR REPORTING OF CONTINUOUS MONITORING DATA

Operator : CBISS Ltd  
Location : 11 Ark Royal Way

Authorisation No : E2242R  
Parameter : SO2 (high scale) mg/m<sup>3</sup>

Release Point : 1  
Reporting Period : September



## Who are CBISS?

- We are an ISO 9001 accredited engineering company based in the North West of England.
- We specialise in Emissions monitoring and gas analysis
- We employ over 50 people and have offices across the UK



## Why do we have CEMS ?

- Environmental - Green and clean !
- Process optimisation !
- **IN THE MAJORITY OF CASES ITS DUE TO LEGISLATION**
- Waste Incineration Directive (WID)
- Large Combustion Plant Directive (LCPD)
- IPPC/IPC

a<sup>1</sup>



monitoring & control for your environment

# MCERTs

Working under either WID or LCPD your permit/licence will ask for MCERTs or TUV approval

The certified range must be a minimum of 1.5 Times the ELV for WID or 2.5 Times for LCPD

E.g.: HCL 0 – 15mg/m<sup>3</sup> on an daily ELV of 10mg/m<sup>3</sup>



## PRODUCT CONFORMITY CERTIFICATE

This is to certify that the

**MIR9000 Multi-gas Analyser  
Type 2 SEC Probe  
Type 2 TIG Calibration/blow back module**

manufactured by:

**Environnement SA**  
111 Boulevard Robespierre  
78304 Poissy Cedex  
France

has been assessed by Sira Certification Service  
and found to comply with:

**MCERTS Performance Standards for Continuous Emission  
Monitoring Systems (November 1998)**

Certification Ranges :

SO <sub>2</sub> 0 - 75 mg/m <sup>3</sup> (l to 25 ppm),	0 - 200 mg/m <sup>3</sup> (0 to 70 ppm)
CO 0 - 75 mg/m <sup>3</sup> (l to 60 ppm),	0 - 250 mg/m <sup>3</sup> (0 to 200 ppm)
NO 0 - 80 mg/m <sup>3</sup> (l to 60 ppm),	0 - 335 mg/m <sup>3</sup> (0 to 250 ppm)
HCl 0 - 15 mg/m <sup>3</sup> (0 to 10 ppm),	0 - 100 mg/m <sup>3</sup> (0 to 60 ppm)

Certification is awarded in respect of the conditions stated in this certificate

Project No: 474/R1/04  
 Certificate No: Sira MC 00010/01  
 Initial Certificate: 10 January 2002  
 This Certificate Issued: 21 February 2005  
 Renewal Date: 02 January 2007

Chief Executive

MCERTS is operated on behalf of the Environment Agency by

**Sira Certification Service**  
South Hill, Chislehurst, Kent BR7 5EH, England  
Tel: 020-8467-2636 Fax: 020-8468 1841

This certificate may only be reproduced in its entirety and without charge

## Heated FID – EN12619

- For WID applications the VOC measurement must utilise a Heated FID instrument.
- This again must be MCERT/TUV approved and the relevant certification range for 1.5 times ELV
- Most permits/licenses will ask for compliance with EN12619 which requires the measurement to be hot/wet
- Currently only heated FID's have met this standard
- Burners placed in a heated furnace, up to 190°C allowing measurement of high concentrations of heavy hydrocarbons



a<sup>1</sup>



monitoring & control for your environment


# Duplex Systems



# Duplex Systems

- External O<sub>2</sub> to offer full redundancy without the requirement to switch the whole system
- Particulate compliant with EN13284
- Sample extraction probe/conditioning unit
- Integrated Temperature, Pressure & Flow
- Heated FID for VOC's
- Live data comparison between duty & standby system on C-DAS


PC  
MIRFT  
GRAPHITE



PC  
MIRFT  
GRAPHITE

SEC BOX

Heated FID




Live Duty  
&  
Standby Data



monitoring & control for your environment

# MIR-FT Example Duplex System





monitoring & control for your environment

# Duplex Screen Shot

CDAS 2004

File Edit View Users Create Window Help

Digital Indicator

**Stream 1 Standby MIR-FT Selected**

Stream 1 Duty Oxygen

Stream 1 Duty Dust

Digital Indicator

**Standby Analyser Selected to Stream One**

Standby Analyser Not Selected to Stream Two

Stream One Duty Analyser To Spare

Digital Indicator

**Stream 2 Duty MIR-FT Selected**

Stream 2 Duty Oxygen

Stream 2 Duty Dust

Stream One	Corrected Data	Adjusted Data	Rolling Ten Minute Average
NOx mg/Nm3	30.21	24.17	24.07
SO2 mg/Nm3	116.60	93.28	92.89
HCl mg/Nm3	5.30	3.18	3.17
CO mg/Nm3	10.60	9.54	9.50
Ammonia mg/Nm3	4.24	2.54	2.54
Active O2 %	11.56	11.56	11.52
Standby O2 %	11.36	11.36	11.36
Active Dust mg/Nm3	6.97	4.88	4.87
Standby Dust mg/Nm3	6.82	4.77	4.76
H2O %	6.84	6.84	6.84
VOC FID mg/Nm3	2.49	1.74	1.74
Flue Gas Flow Nm3/hr	2546.00	2546.00	2546.00
Flue Gas Temp oC	256.00	256.00	256.00
Flue Gas Pressure mBar	N/A	N/A	N/A

Spare Analyser	Corrected Data	Adjusted Data	Rolling Ten Minute Average
NOx mg/Nm3	30.21	24.17	24.05
SO2 mg/Nm3	116.60	93.28	92.80
HCl mg/Nm3	5.30	3.18	3.16
CO mg/Nm3	10.60	9.54	9.49
Ammonia mg/Nm3	4.24	2.54	2.53
Active O2 %	11.56	11.56	11.51
Standby O2 %	11.36	11.36	11.36
Active Dust mg/Nm3	6.97	4.88	4.86
Standby Dust mg/Nm3	6.82	4.77	4.75
H2O %	6.84	6.84	6.84
VOC FID mg/Nm3	2.49	1.74	1.74
Flue Gas Flow Nm3/hr	2546.00	2546.00	2546.00
Flue Gas Temp oC	256.00	256.00	256.00
Flue Gas Pressure mBar	N/A	N/A	N/A

Stream Two	Corrected Data	Adjusted Data	Rolling Ten Minute Average
NOx mg/Nm3	29.21	23.37	24.09
SO2 mg/Nm3	112.73	90.19	92.98
HCl mg/Nm3	5.12	3.07	3.17
CO mg/Nm3	10.25	9.22	9.51
Ammonia mg/Nm3	4.10	2.46	2.54
Active O2 %	11.24	11.24	11.53
Standby O2 %	11.36	11.36	11.36
Active Dust mg/Nm3	6.74	4.72	4.87
Standby Dust mg/Nm3	6.59	4.61	4.76
H2O %	6.84	6.84	6.84
VOC FID mg/Nm3	2.41	1.69	1.74
Flue Gas Flow Nm3/hr	2546.00	2546.00	2546.00
Flue Gas Temp oC	256.00	256.00	256.00
Flue Gas Pressure mBar	N/A	N/A	N/A

Corrected Data Received from West Stack 22 September 2004 14:41:27

CBISS Software Engineer

14:41



monitoring & control for your environment

# Data Acquisition & Reporting






monitoring & control for your environment

# C-DAS Data Acquisition Software

- Corrects values for flow, pressure, temperature
- Reports mass emissions
- Allows for plant run time
- Generates alarms
- 8 years 100+licenses
- Combustion, boilers, incineration, chemical, Cement, etc





monitoring & control for your environment

# Information Grids (configurable)

Channel 1	Date	Time	Raw	Corrected	Daily Average	Max 10 Min TWA	Max 30 Min TWA	Daily Average Hi Limit	Exceeds of Daily Average Hi Limit
O2 %	23/08/2002	10:10:57	12.1	N/A	12.0	12.2	12.2	18.0	0
CO mg/m3	23/08/2002	10:10:57	249.80	281.03	305.83	410.71	410.71	400.00	0
HCl mg/m3	23/08/2002	10:10:57	721.9	812.1	452.3	647.2	647.2	100.0	3
TOC mg/m3	23/08/2002	10:10:57	203	229	241	291	252	500	0
NO mg/m3	23/08/2002	10:10:57	11.10	12.49	13.33	19.33	19.33	20.00	0
NH3 mg/m3	23/08/2002	10:10:57	24.9	28.0	22.4	29.9	22.4	30.0	0
SO2 mg/m3	23/08/2002	10:10:57	19.8	22.3	15.2	20.2	20.2	100.0	0
NO2 mg/m3	23/08/2002	10:10:57	47.30	53.21	54.94	89.33	89.33	120.00	0
HF mg/m3	23/08/2002	10:10:57	42.0	47.3	51.2	64.1	64.1	120.0	0
Temperature C	23/08/2002	10:10:57	607	N/A	696	777	777	1000	0
Flow Rate m3/h	23/08/2002	10:10:57	25055.60	N/A	24264.85	26065.05	26065.05	40000.00	0

CDAS 2002  
File Edit View Create Window Help


LATEST : 2,16/05/2002,16:01:09,201.000,203.000,201.000,198.000,154.100,201.000,2953.000,22

Incinerator 1	Date	Time	Raw	Corrected	Last 1 Min Avg	Last 10 Min Avg	Last 30 Min Avg	10 Min TWA	30 Min TWA	60 Min TWA
HCl mg/m3	6/05/2002	16:00:59	216.0	216.0	215.0	211.3	209.5	211.8	209.9	209.9
SOS2 mg/m3	6/05/2002	16:00:59	196.0	196.0	195.0	203.9	202.7	203.3	202.2	202.2
NO2 mg/m3	6/05/2002	16:00:59	206.0	206.0	203.0	198.9	199.4	199.1	199.6	199.6
CO mg/m3	6/05/2002	16:00:59	211.0	211.0	209.5	202.3	201.3	203.2	201.8	201.8
CO2 %	6/05/2002	16:00:59	152.1	N/A	151.0	149.1	148.9	149.5	149.1	149.1
HIC mg/m3	6/05/2002	16:00:59	200	200	190	190	190	190	190	190
H2O ppm	6/05/2002	16:00:59	2794.0	N/A	2811.5	2862.1	2896.8	2856.3	2891.1	2891.1
O2(D) %	6/05/2002	16:00:59	22.00	N/A	22.00	21.85	21.89	21.85	21.90	21.90
Dust mg/m3	6/05/2002	16:00:59	39.3	N/A	56.9	58.5	48.5	58.3	49.1	49.1



monitoring & control for your environment



# On-Line Trends



## ANNEX 2 RELEASES INTO AIR REPORTING OF CONTINUOUS MONITORING DATA

Authorisation No : E2212R  
Parameter : SO2 (high scale) mg/m3

Release Point :A1  
Reporting Period : September 2004




# Corrections and Re-transmissions

Other inputs

Alarms inputs

Scaling Factors

Analogue outputs



**Analogue Input Signals**

Channel	mAV	Scaled
Inc 1 Dust mg/m3	11.559	47.244
Inc 1 Flow k m3/hr	7.707	27.802
Inc 1 Temperature deg C	9.807	108.881
Inc 1 O2 Wet %	14.871	16.986
Inc 2 Dust mg/m3	16.592	78.700
Inc 2 Flow k m3/hr	15.688	87.660
Inc 2 Temperature deg C	14.249	192.169
Inc 2 O2 Wet %	16.567	19.636
Inc 3 Dust mg/m3	4.439	2.744
Inc 3 Flow k m3/hr	7.343	25.073
Inc 3 Temperature deg C	6.076	38.925
Inc 3 O2 Wet %	10.543	10.223

Hide

**Digital Input Signals**

Input	Status
Inc 1 Plant Run	Green
Inc 2 Plant Run	Green
Inc 3 Plant Run	Green
Analyser Fault	Green
Calibration in Progress	Green
Spare	Red
Last Spare	Red

Hide

**Analogue Output Signals**

Inc 1 Dust mg/m3	14.56
Inc 1 Flow k m3/hr	4.00
Inc 1 Temp deg C	4.00
Inc 2 Dust mg/m3	6.32
Inc 2 Flow k m3/hr	4.00
Inc 2 Temp deg C	4.00
Inc 3 Dust mg/m3	5.81
Inc 3 Flow k m3/hr	4.00
Inc 3 Temp deg C	4.00


Hide



monitoring & control for your environment

# Reporting Software

- Off-line Reporting Software
- Uses daily log files
- Produces monthly, quarterly etc. reports
- Easy to use calendar
- Haz waste variation displays
- Customised reports available
- Supported in the UK & Ireland





monitoring & control for your environment

# Daily Software Reports

- Shift Reports
- Daily Log of Raw data
- User defined text “incinerator 1”
- Daily Report
- On demand Report / Printout
- Print on Alarm
- IPPC Annexes

**CBISS DEMO SYSTEM Emission Report**  
From 15 May 2002 00:00 to 15 May 2002 23:59  
Report printed on 15/05/2002 at 15:33:43

Incinerator 1	Corrected NO mg/m <sup>3</sup>	Corrected NO <sub>x</sub> mg/m <sup>3</sup>	Corrected SO <sub>2</sub> Erg mg/m <sup>3</sup>	Corrected HCl mg/m <sup>3</sup>	Corrected CO mg/m <sup>3</sup>
Maximum Value	345.7	344.2	276.2	263.0	318.8
Average Value	312.2	314.5	250.9	158.0	282.9
No. Alarms	0	0	0	0	0
Minimum Value	171.3	170.4	137.7	103.3	163.4
Maximum Hourly Avg	312.2	314.5	250.9	158.0	282.9


Incinerator 1	Corrected -1F mg/m <sup>3</sup>	Raw O <sub>2</sub> %	Corrected TOC mg/m <sup>3</sup>
Maximum Value	344.3	15.28	240.8
Average Value	311.4	13.28	218.1
No. Alarms	0	0	0
Minimum Value	173.9	7.28	119.9
Maximum Hourly Avg	311.4	13.28	218.1

Incinerator 2	Corrected NO mg/m <sup>3</sup>	Corrected NO <sub>x</sub> mg/m <sup>3</sup>	Corrected SO <sub>2</sub> Erg mg/m <sup>3</sup>	Corrected HCl mg/m <sup>3</sup>	Corrected CO mg/m <sup>3</sup>
Maximum Value	329.6	327.7	259.1	261.2	301.0
Average Value	322.7	323.7	250.3	156.3	291.3
No. Alarms	0	0	0	0	0
Minimum Value	327.8	319.4	250.3	157.2	282.3
Maximum Hourly Avg	322.7	323.7	250.3	156.3	291.3

Incinerator 2	Corrected -1F mg/m <sup>3</sup>	Raw O <sub>2</sub> %	Corrected TOC mg/m <sup>3</sup>
Maximum Value	330.8	15.11	237.5
Average Value	325.3	14.23	230.9
No. Alarms	0	0	0
Minimum Value	310.8	14.55	224.0
Maximum Hourly Avg	325.3	14.23	230.9

CBISS DEMO Daily Report for Streams Over 100 Tons 2002 Incinerator 2002  
Report Printed On 15/05/2002 On Tue 15 December 2002

Stream	NO	NO <sub>x</sub>	SO <sub>2</sub>	HCl	CO	-1F	O <sub>2</sub>	TOC
10000	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10001	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10002	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10003	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10004	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10005	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10006	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10007	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10008	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10009	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10010	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10011	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10012	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10013	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10014	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10015	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10016	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10017	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10018	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10019	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10020	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10021	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10022	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10023	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10024	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10025	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10026	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10027	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10028	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10029	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10030	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10031	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10032	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10033	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10034	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10035	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10036	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10037	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10038	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10039	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10040	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10041	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10042	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10043	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10044	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10045	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10046	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10047	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10048	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10049	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10050	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10051	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10052	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10053	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10054	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10055	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10056	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10057	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10058	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10059	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10060	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10061	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10062	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10063	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10064	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10065	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10066	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10067	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10068	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10069	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10070	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10071	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10072	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10073	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10074	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10075	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10076	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10077	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10078	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10079	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10080	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10081	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10082	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10083	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10084	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10085	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10086	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10087	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10088	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10089	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10090	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10091	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10092	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10093	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10094	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10095	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10096	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10097	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10098	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10099	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1
10100	312.2	314.5	250.9	158.0	282.9	311.4	13.28	218.1





monitoring & control for your environment



# Service & Support

**Paul Burgess**  
(Service Manager)



**Adam Dabrowski**  
(Regional Service Manager - North)



**Paul Morgan**  
(Regional Service Manager - South)



**John Dakers**  
(Senior Service Engineer - Scotland)



**Lee Spencer**  
(Senior Service Engineer - South)



**Dee Lound**  
(Service Coordinator & Supervisor)



**Iain McGlinchey**  
(Service Engineer - Scotland)



**Daryl Wilson**  
(Service Engineer - North)



**Stuart Barrowcliff**  
(Service Engineer - North)



**John Hall**  
(Service Engineer - North)



**Paul Reeve**  
(Service Engineer - South)



**Andy Dowling**  
(Service Engineer - South)



**Marc Roach**  
(Bench Service Engineer)



**Angela Evans**  
(Service Admin)



monitoring & control for your environment

## **Service & Support**

24 hour 7 days per week cover

Direct engineer contact day or night

14 person service team

11 Factory trained dedicated engineers

Remote system diagnostics through modem

Software & IT support for supplied equipment

Commissioning to relevant performance standards



monitoring & control for your environment

## Commissioning

Full function check of complete system including standard zero and span calibration of the analyser.

Verification of compliance to relevant performance standards.

NOX - ISO10849

CO,CO<sub>2</sub> and O<sub>2</sub> - ISO12039

SO<sub>2</sub> - BS6069

# EN 14181 – The QAL's

## Quality Assurance of Automated Measuring Systems

*“(QAL 1) to demonstrate that the AMS is suitable for the intended purpose before installation, by meeting required performance standards and the uncertainty budgets specified in the EU directives;”*

*“(QAL 2) to calibrate the AMS and determine the variability of the measured values obtained by it; so as to demonstrate the suitability of the AMS for its application, following installation;”*

*“(QAL 3) to maintain and demonstrate the required quality of the measurement results during the normal operation of the AMS, by checking that the zero and span characteristics are consistent with those determined during QAL 1;”*



monitoring & control for your environment

# EN 14181 – QAL 1

## Quality Assurance of Automated Measuring Systems

*“(QAL 1) to demonstrate that the AMS is suitable for the intended purpose before installation, by meeting required performance standards and the uncertainty budgets specified in the EU directives;”*

# EN 14181 – QAL 1

## SUMMARY REPORT - Waste Incineration

Expanded Uncertainty of ENVIRONNEMENT SA's AMS : MIR 9000, MIR FT and GRAPHITE 52M



AMS	Gas	Expanded uncertainty at 95% (in mg/m <sup>3</sup> dry without O <sub>2</sub> correction)	Expanded uncertainty at 95% (in mg/m <sup>3</sup> dry corrected @11%O <sub>2</sub> )	Expanded uncertainty at 95% (in % of ELV dry corrected @11%O <sub>2</sub> )	Maximum Authorized Expanded Uncertainty according to Directive 2000/76/CE	AMS acceptable (Yes/No)
MIR 9000	HCl	3,38	3,40	34,0%	40,0%	Yes
MIR 9000	SO <sub>2</sub>	6,73	6,07	13,5%	20,0%	Yes
MIR 9000	NO	7,49	10,48	3,7%	20,0%	Yes
MIR 9000	NO <sub>x</sub>	7,89	11,03	3,9%	20,0%	Yes
MIR 9000	CO	2,83	3,37	5,7%	10,0%	Yes
MIR 9000	TOC	under evaluation	under evaluation	under evaluation	30,0%	-
GRAPHITE 52M (used with MIR 9000)	TOC	0,83	0,81	9,1%	30,0%	Yes
MIR FT	HCl	0,74	0,84	8,4%	40,0%	Yes
MIR FT	SO <sub>2</sub>	2,79	3,40	6,8%	20,0%	Yes
MIR FT	NO	11,82	16,74	5,5%	20,0%	Yes
MIR FT	NO <sub>x</sub>	12,44	17,62	5,8%	20,0%	Yes
MIR FT	CO	2,51	3,17	6,3%	10,0%	Yes
MIR FT	TOC	0,83	1,00	10,0%	30,0%	Yes
GRAPHITE 52M (used with MIR FT)	TOC	0,86	0,94	9,4%	30,0%	Yes

**Nota :**

- 1) NO & NO<sub>x</sub> expressed as mg/m<sup>3</sup> eq NO<sub>2</sub>. TOC expressed as mgC/m<sup>3</sup>
- 2) Data from TÜV reports for MIR FT and GRAPHITE 52M are on a dry basis
- 3) Data from MCERT report MIR 9000 and MIR 9000 O<sub>2</sub> sensor are on dry basis
- 4) Data for MIR FT CO sensor are on wet basis

# EN 14181 – QAL 1

## Expanded Uncertainty of MIR 9000 AMS for use in Waste Incineration in the EU

Gas : HCl

MIR 9000

Certification Range (CR) : 98,4 mg/m<sup>3</sup> (MCERT / 60 ppm)  
 Test concentration (TC) : 82,0 mg/m<sup>3</sup> (MCERT / 50 ppm)  
 Emission limit value (ELV) : 10,0 mg/m<sup>3</sup> corrected to dry @11%vO<sub>2</sub> (daily)



Partial uncertainty	Result of MCERT test			u (HCl) (mg/m <sup>3</sup> )	u <sup>2</sup> (HCl)	Remarks
	% of CR or TC	mg/m <sup>3</sup> referred to CR or TC	mg/m <sup>3</sup> referred to ELV			
Refer to "Formula" sheet. Equation n° x						
n°1 - Uncertainty due to linearity (lack of fit) : u <sub>lin</sub>	0,710%	0,699	-	0,403	0,1627	MCERT Laboratory Evaluation Report page 18-Table 6a
n°2 - Uncertainty due to zero drift : u <sub>zdrift</sub> (zero drift)	-	-	-	-	-	Zero drift < Span drift : not considered
n°3 - Uncertainty due to span drift : u <sub>zdrift</sub> (span drift)	1,700%	1,394	0,170	0,098	0,0098	MCERT Field Evaluation Report - Page 27 (% TC / week)
n°4 - Uncertainty relating from variations in ambient pressure expressed as standard deviation : u <sub>press</sub>	N.A.	N.A.	N.A.	0,000	0,0000	N.A. : no sensitivity to atmospheric pressure variations (compensation)
n°5 - Uncertainty relating from variations in ambient temperature expressed as standard deviation : u <sub>temp</sub>	2,366%	2,331	-	1,346	1,8105	MCERT Laboratory Evaluation Report page 30-Table 10a (interpolation at 20°C +/-5°C)
n°6 - Uncertainty due to cross interferences : u <sub>zdrift</sub> (int)	-	-1,582	-	-0,913	0,8343	MCERT Laboratory Evaluation Report page 22-Table 7a. Max (Σ interferences +, Σ interferences -)
n°7 - Standard deviation of repeatability at zero : S <sub>r,0</sub>	0,045%	0,045	-	0,026	0,0007	MCERT Laboratory Evaluation Report page 26-Table 8a. S <sub>r,0</sub> < S <sub>r,s</sub> : not considered
n°8 - Standard deviation of repeatability at test concentration : S <sub>r,s</sub>	0,177%	0,175	-	0,175	0,0305	p 27 tab 9a repeatability divided by t student ( 2,65)
n°9 - Uncertainty due to span gas unaccuracy : u <sub>span</sub>	-	2,000%	0,20	0,115	0,0133	span gas relative accuracy : 2%
n°10 - Uncertainty relating from correction to dry : u <sub>condry</sub>	N.A.	N.A.	N.A.	N.A.	N.A.	Measurements are made on a dry sample thanks to SEC sampling device (permeation-based)
					2,8610	Sum of u <sup>2</sup> (HCl), without O <sub>2</sub> correction
n°11 - Uncertainty relating from influence of correction to O <sub>2</sub> : u <sub>vO2</sub>					0,0335	Refer to "O <sub>2</sub> " uncertainty calculation sheet
					2,8945	Sum of u <sup>2</sup> (HCl), with O <sub>2</sub> correction



Combined uncertainty u <sub>c</sub> (HCl) :	1,691 mg/m <sup>3</sup> (dry without O <sub>2</sub> correction)	$\sqrt{\sum u^2(HCl)}$
Combined uncertainty u <sub>c</sub> (HCl) :	1,701 mg/m <sup>3</sup> (dry with O <sub>2</sub> correction)	
Expanded uncertainty U(HCl) at 95% confidence interval (K=2) :	3,38 mg /m <sup>3</sup> (dry without O <sub>2</sub> correction)	
Expanded uncertainty U(HCl) at 95% confidence interval (K=2) :	3,40 mg /m <sup>3</sup> (dry at 11%O <sub>2</sub> )	
Expanded uncertainty U(HCl) at 95% confidence interval (K=2) :	34,0% of ELV (dry @11%O <sub>2</sub> )	
Required U <sub>req</sub> (HCl) according to 2000/76/CE Directive :	40,0% of ELV (dry @11%O <sub>2</sub> )	
<b>U(HCl) &lt; U<sub>req</sub>(HCl) - MIR 9000 for HCl complies with Waste Incineration Directive 2000/76/CE</b>		

Calculations made according to IEN ISO 14956 Standard - MIR HCl data - v8, December 9th, 2004 - DM/SA



monitoring & control for your environment

# EN 14181 – QAL 2

## Quality Assurance of Automated Measuring Systems

*“(QAL 2) to calibrate the AMS and determine the variability of the measured values obtained by it; so as to demonstrate the suitability of the AMS for its application, following installation;”*


## EN 14181 – QAL 2



- Instrument Compared to Standard Reference Method (SRM) in a Parallel Test.
- Results From Two Instruments Compared and Linear Calibration Function Generated
- All Data from the AMS is now Calibrated in Software Against This Function.
- Range of Data used to Generate Function Provides Valid Instrument Range. Emissions Should Remain Within This Range or New QAL 2 Performed.

Appendix 1 - Raw Data, Calculations & Graphs

Hydrogen Chloride - x-y Plot of Parallel Measurements





## **Traditionally**

- End users would usually have existing relationships with stack testing houses
- The same companies who carry out periodic testing would be approached for the QAL 2 test
- Site specific experience, existing account info, trusted engineers & fully inducted etc
- As long as the company has the correct certification of personnel, equipment and procedure then, this option is not always a bad one.



# **The Danger Of Separation**

**It may seem sensible for an operator to Order a AMS systems from a AMS supplier, and order a Stack test from a Stack tester.**

**This can and has caused problems in the past eg**

- Results From Stack test compared to wrong data**
- Results from stack tests not corrected in the same way as CEMS , (moisture O2 etc)**
- Analyser (s) offline/in fault for part of tests**
- Analyser Calibration / zero ref not adjusted**
- Test carried out when analyser is due for calibration / service**
- Usually results in tests having to be repeated at a cost and contractual disputes**



monitoring & control for your environment

## AMS Provider Involved In QAL 2

Due to our previous experience with stack testing we recommend to our customers that we are **involved** in all QAL 2 testing.



The majority of systems we are installing today include the QAL 2 test as our scope of supply. As an example CBISS hired an qualified Test House to do this work with our assistance.

On reviewing the Standard we found that it was essential that the Stack tester and CBISS understood the standard, and the AMS equipment.

# Working Relationship

- Initial meeting to discuss Standard and equipment tests
- Produce test document specific to equipment and site
- Stack tester is trained to be able to view analyser error messages make simple analyser checks and be able to take data from AMS reporting software
- CBISS perform service & calibration visit on site prior to Stack tests
- CBISS engineer & Stack tester visit site for first day of testing
- Report completed and passed to customer.

Usually results in tests not repeated.



## QAL 2 Report

- Once the tests have taken place and the successful report has been issued, the data is utilised by the software package
- The AMS must be calibrated to the SRM results
- Various software packages are available with an additional upgrade specifically for the EN14181 standard





monitoring & control for your environment

# Example QAL 2 software

CDAS 2004 DATA COLLECTION MODULE

File View Edit Help

Demo MIR

	Tag Description	Instrument		EN 14181		Calibrated Value
		Value	Scale	Gradient	Intercept	
Gas 1	HCl	8.8	xxx.x	1	0	8.8
Gas 2	SO2	23.7	xxx.x	1	0	23.7
Gas 3	NO	0	xxx.x	1	0	0
Gas 4	NO2	0	xxx.x	1	0	0
Gas 5	CO	6.4	xxx.x	1	0	6.4
Gas 6	HF	0	xxx.x	1	0	0
Gas 7	H2O	9582	xxxx	1	0	9582
Gas 8	CO2	0	xxx.x	1	0	0
Gas 9	O2	22	xxx.x	1	0	22
Gas 10	Item 10	0	xxxx	1	0	0
Gas 11	Item 11	3000	xxxx	1	0	3000
Gas 12	Item 12	4000	xxxx	1	0	4000
Gas 13	Item 13	N/A	xxxx	1	0	N/A
Gas 14	Item 14	N/A	xxxx	1	0	N/A

Get Up Gas Values Instrument Status



monitoring & control for your environment

# EN 14181 – QAL 3

## Quality Assurance of Automated Measuring Systems

*“(QAL 3) to maintain and demonstrate the required quality of the measurement results during the normal operation of the AMS, by checking that the zero and span characteristics are consistent with those determined during QAL 1;”*

**This is the end users responsibility to carry these tests out**

## **EN 14181 – QAL 3**

- Regular zero and span check to ensure equipment continues to meet the uncertainty specification.
- Results of these measurements need to be recorded.
- Analysis of the all recent recorded results must be performed.
- If necessary actions must be taken if results of analysis show problems with performance of the monitoring system.

## **EN14181 - Analysis of Results**

- EN 14181 QAL 3 allows for two methods of analysing results.
  - Shewhart Chart
  - CUSUM
- A report will be generated upon completion of the QAL 3 calibration and any problems arising should be clearly noted on the report.
- Possible to recreate old reports as well as viewing results of the calibration.

## **EN 14181 – QAL 3**

**Manual QAL 3 procedure would consist of:**

- **Injecting each calibration gas either to the probe or analyser**
- **Logging of the data either manually or from the software package**
- **Comparing of the data against the relevant calibration gas values**
- **Calculate Shewhart/CUSUM control charts**
- **QAL 3 report**

## QAL 3 Calibration checks

**Configure EN 14181 QAL 3 Calibration**

Configure Calibration Periods

Period of Calibration:

Select the Day of the Week:

Time Of Day:

User Confirmation:

Manual Calibration

Selecting the button below will cause the software to begin a QAL 3 calibration check on all instruments. Only Select this option if you are sure you wish to perform this procedure.

**QAL 3 Calibration Alert**

**A QAL 3 Calibration is Due. Click OK to Perform this Calibration Now or Click Cancel to Abort this Calibration**


# Example of a CUSUM Report

Keep it simple

Keep the level of input from the operators to a minimum

The easiest way to control and comply with EN14181 is through software

**Weekly QAL 3 Report  
Gosford Incinerator  
CUSUM Chart for CO : Analyser MIR9000  
Report Generated 10<sup>th</sup> January 2005  
Report for Data from 2/9/04 to 5/1/05**



Summary of Analysis

$$S_{ams} = 0.789$$

### Drift

$$S_t > 2.850 \quad S_{ams} : 2.850 > 2.249 \quad \text{OK}$$

$$S_t > 0.501 \quad S_{ams} : 0.254 > 0.395 \quad \text{OK}$$


### Precision

$$S_t > 6.900 \quad S_{ams} : 3.900 > 5.444 \quad \text{OK}$$

$$S_t > 1.850 \quad S_{ams} : 1.150 > 1.460 \quad \text{OK}$$


## EN 14181 – QAL 3

- When a QAL 2 is completed the AMS should be QAL 3 “ready”.
- Analysers need to carry out calibration checks against span gas and zero readings at regular intervals.
- Results of these measurements need to be recorded.
- Analysis of all recently recorded results must be performed.



**QAL 3 REPORTS**  
Cusum or Shewhart

**Data logging & Control of  
Gas injection**



**IR analyser**  
**Cal Gas BOX**  
**FID Analyser**

## EN 14181 – AST

- **AST - Annual Surveillance Test** used to ensure QAL 2 calibration function is maintained. Consists of alignment/cleaning of sampling system, leak test, zero/span, linearity, response time, interference check **plus 5 parallel measurements** spread evenly over one day, checking documentation and records followed by issuing of a report



## Summary

- Find out all of your stack conditions - temp, moisture, flow, particulate....
- Select the best technique for your environment – In-situ, Extractive....
- Compliant with relevant standards – WID, LCPD, MCERTS...
- Select a company who have the backup for all integrated products
- Select a software package which is flexible
- Turnkey contract from design to commissioning
- Remote diagnostics if possible
- Duplex systems - Justifiable?