



Practical Experiences of WID Process Installations to Meet EN-14181



Effective Procedure for EN-14181

- A 3 Step approach
 - QAL1; Ensure appropriate approval including certification range (include sufficient pre order analysis)
 - QAL3; Use of AST Audit Units and self check results
 - QAL2; Functionality checks prior to 13284-1. Also discussions with plant and Iso team to see what measures can be done re raising/reducing emissions




Effective Procedure for EN-14181

- Examples of typical installations
 - 2 DT991 applications (Electrodynamic Probe)
 - 2 LMS181 applications (Proscatter)Including some QAL2 reporting



Installation Considerations

- Give priority to locating the Isokinetic sampling ports
- There are specific guidelines as to where the Iso ports should be located
- The sensor should be located >2 duct diameters downstream of Iso ports – This is a guide. It must be clarified by means of additional tests
- Care on orientation when installing instrument 'mounting' to minimise interference



Installation Considerations

2.1.3 Criteria for locating the sampling plane (Extracted from M1, with view to draft prEN15259)

The general approach for periodic sampling of particulates and aerosols can be summarised as:

Sampling must be carried out at a suitable location on the stack. Bends, branches, obstructions, fans and leaks can all cause undesirable variations in the velocity profiles, which may make the location unsuitable for sampling.

For new installations, the sampling plane should be located according to the requirements contained in Table 2.1. By adhering to these requirements, there is a very good chance that the flow-stability criteria defined in Table 2.2 will be met. The flow-stability criteria of a gas encompass velocity and direction. Flow criteria is used in this TGN as a term to cover all these parameters. It should be emphasised that it is the flow criteria that are important, not the adherence to the positional requirements *per se*. **Therefore on plants where the sampling plane location does not meet the recommendations in Table 2.1, it is possible to locate the sampling plane at another location as long as the flow criteria are met.**


Identify a potentially suitable sampling location (covered in this section) Access the stack and check that the gas-flow criteria are met by carrying out an exploratory survey (see Sections 2.1.4 and 2.1.5) Decide the number and position of the sampling points (see Sections 2.1.7 and 2.1.8), and install ports and access to the sampling location (see Section 4.1).

Carry out preliminary velocity traverse (Section 2.1.5) to confirm satisfactory and proceed with sampling.

Installation Considerations


- If it is not possible to mount in ideal location, then install Dust monitor in best location in relation to the Iso ports, bearing in mind the need to understand implications of 'obstacles'

- Ensure a Pitot traverse is carried out to confirm flow profile – including possible affects of Iso equipment




Check orientation and distance from stack wall

Ensure sufficient clearance in ducting
NB Same stack mounting for DT991 and LMS181



Flange Information

Dual Flange arrangement




DIMENSION	3" 150lb ANSI	DN80 PN10 DN80 PN 16
OD	190mm	200mm
PCD	152.4mm	160mm
HOLE SIZE	19mm	18mm

DIMENSION	3" 150lb ANSI	DN80 PN 6	DN80 PN10 DN80 PN 16
OD	190mm	190mm	200mm
PCD	152.4mm	150mm	160mm
HOLE SIZE	19mm	18mm	18mm


Application 1, DT991 MWI

Stack Split in to two parts by means of dividing plate running the length of the stack



Application 1, DT991 MWI

Flange
Face




Insulation thickness.
Mounting 'tube' through
into the duct from Flange



Application 1, DT991 MWI




Application 1, DT991 MWI



Application 1, DT991 MWI




Water Build
up on Upward
Pointing
Mounting



Water




Carbon
Build up
all points



Dust 3, 1st Time


The DT991 Sensors

Standard



991 for large stacks and stacks with 'stand offs' – typical if replacing existing optical system

Option




991 for small ducts

Both have option for higher temperature of 400°C


DT991 System

System Overview



DT991 Mechanical Arrangements

Standard Passive Active Option –
With Air Purge




Blanking Plug
removed and Air
Coupling added

Blanking Plugs
removed and Air
Purge Body Added

DT991 Mechanical Arrangements

Standard Passive Active Option –
With Air Purge



Blanking Plug
removed and Air
Coupling added

Blanking Plugs
removed and Air
Purge Body Added



Why DT991 has Passive Active as standard


Flue gas Temperature
above dew point

Condensation
around base as
temperature drops
below dew point

OPTIONS –


Air Purge

Passive Active



Why DT991 has Passive Active as standard

If long 'Stand Off' 150mm+ eg through insulation, then water can collect in the 'tube' which can cause interference signal –




Solution is to use Passive / Active Style



Why DT991 has Passive Active as standard

If long 'Stand Off' 150mm+ eg through insulation, then water can collect in the 'tube' which can cause interference signal –



Solution is to use Passive / Active Style



Installation Considerations




Examples of how
not to do it




Application 1, DT991 MWI

- Long term log showing good tracking when originally fitted, problems encountered and finally corrected




Application 1, DT991 MWI

- Both DT991 sensors tracking well – In this case CU1, CH2 (Green) CU2, CH1 (Red)



Application 1, DT991 MWI

- Both DT991 sensors tracking well
 - In this case CU1, CH1 (Green)
 - CU2, CH2 (Red)



Application 2, DT991

Electrodynamics




Application 2, DT991

- AST Unit used prior to installation




Application 2, DT991


- AST Unit used prior to installation



Application 2, DT991




Application 2, DT991




Application 3, LMS181 Proscatter




Application 3, LMS181



Application 3, LMS181




Application 3, LMS181



Application 3, LMS181

Additional tests
for instruments
zero –
Achieved by
removing
sensor from
stack and use
of 'surrogate'
(as indicated in
M20)



Application 4, LMS181 Co-Incinerator




Application 4, LMS181 Co-Incinerator



Certification

Application 4, LMS181 Co-Incinerator




- CH 1 (Green) & CH2 (Red) Combined to show how well they tracked each other



Application 4, LMS181 Co-Incinerator



This shows a nice cluster.

In this case, follow the guidelines in M20 – average the 15 points, then draw a line through zero and extend upward (by 100%).



Application 4, LMS181 Co-Incinerator


Again, a nice cluster with appropriate implementation of M20 guidelines.



Example of QAL3 data - Zero

Shewart Control Chart : Zero Reference (following M20 technical guidance note to EN14181)


- There are the two methods of displaying QAL3 data:
 - 1. Manual using spread sheet template.
 - 2. Automatic download to Dust Reporter
- This is an example of how the data stored can be presented using Shewart formula in an Excel Spreadsheet



Example of QAL3 data - Span

Shewart Control Chart : Span Reference (following M20 technical guidance note to EN14181)

- This is an example of how the data stored can be presented using Shewart formula in an Excel Spreadsheet




Example of QAL3 data - Span

- This is an example of how the QAL3 data stored is presented using QAL3 Software
- Our guide for where to set the 'limits' are;

- First (warning) limit set after reviewing 'instrument drift' over minimum of 7 day period

- Second is limit as directed by directive (30%)



Summary

- Take care with location and installation
- Take time to analyse 'Typical Emissions' prior to arranging QAL2
- Remember QAL2 has 2 parts namely, FUNCTIONALITY and SRM 13284-1
- It will cost you, in more ways than one, if you do not take time to understand your full requirements